

Main 1st Dental



| System Type = Hillrom CWS Unit Main 1st DIALYSIS |                    |     |                                                |
|--------------------------------------------------|--------------------|-----|------------------------------------------------|
| 5/16/2024                                        | PNumber            | QTY | Description                                    |
| NS                                               | AC19               | 1   | INSTALLATION Single-Gang Rough-in              |
|                                                  | P2509VNC300        | 1   | GRS10H DSK MNT. 4USB                           |
| NS                                               | P2505NNC1C00       | 11  | AUDIO STATION BED CONNECTOR                    |
|                                                  | P2505NNC3A02-G     | 12  | ZONE LIGHT, WHIT-TAX COMPLIANT                 |
| NS                                               | P2505NNC3A02       | 1   | ICON BASED-LIGHT LED ZONE LIGHT                |
|                                                  | P2510AC3115-105018 | 10  | PULL SPR. DAMCO DVX 2.17.5                     |
| POE-24                                           | P2519NNC2A24       | 1   | POE SWITCH NNC 24-PORT DELTA II                |
|                                                  | P2519NNCFA00       | 2   | MULTI-MODE G8RC FOR TEL-ROOM POE               |
| NS                                               | P2520NNC3A07       | 1   | PB SWITCH, CODE BLUE, WICANCEL                 |
|                                                  | P2520NNC3B01       | 2   | PULL SWITCH BATH WICANCEL                      |
| NS                                               | P2521B02           | 1   | UPS,RACKMOUNTABLE, 2U - NON-SEISMIC            |
|                                                  | P2594VNC1D01       | 3   | STAFF STATION - STANDARD ROOM STATION W/O CODE |
| NS                                               | P2594VNC1D11       | 12  | PATIENT STATION - STANDARD ROOM STATION W/CODE |
|                                                  | P2599NNC3C00       | 11  | RCB3 ROOM CONTROL BOARD                        |



| System Type = Hillrom CWS Unit Main 1st FL ED |                     |     |                                                |
|-----------------------------------------------|---------------------|-----|------------------------------------------------|
| 5/16/2024                                     | PNumber             | QTY | Description                                    |
| NS                                            | AC19                | 2   | GRS10H DSK MNT. 4USB                           |
|                                               | P2509VNC300         | 27  | ZONE LIGHT, WHIT-TAX COMPLIANT                 |
| NS                                            | P2511BCC-0101-08B-U | 21  | CALL CORD, OXYGEN SAFE, 8 FT                   |
|                                               | P2516A01-G          | 21  | EQUIPMENT RECEPTACLE (TAA)                     |
| POE-24                                        | P2519NNC2A24        | 2   | POE SWITCH NNC 24-PORT DELTA II                |
|                                               | P2519NNCFA00        | 2   | MULTI-MODE G8RC FOR TEL-ROOM POE               |
| NS                                            | P2520NNC3A07        | 1   | PB SWITCH, CODE BLUE, WICANCEL                 |
|                                               | P2520NNC3B01        | 5   | PULL SWITCH BATH WICANCEL                      |
| NS                                            | P2520NNC3B02        | 3   | PULL SWITCH, BATH, W/O CANCEL                  |
|                                               | P2520NNCBH00        | 4   | BH PB SWITCH, WICANCEL                         |
| NS                                            | P2521B02            | 1   | UPS,RACKMOUNTABLE, 2U - NON-SEISMIC            |
|                                               | P2594VNC1D11        | 21  | PATIENT STATION - STANDARD ROOM STATION W/CODE |
| NS                                            | P2599NNC3C00        | 16  | RCB3 ROOM CONTROL BOARD                        |

| System Type = Hillrom CWS Unit Main 1st Oncology |                     |     |                                                |
|--------------------------------------------------|---------------------|-----|------------------------------------------------|
| 5/16/2024                                        | PNumber             | QTY | Description                                    |
| NS                                               | AC19                | 10  | INSTALLATION Single-Gang Rough-in              |
|                                                  | P2509VNC300         | 1   | GRS10H DSK MNT. 4USB                           |
| NS                                               | P2505NNC3A02-G      | 25  | ZONE LIGHT, WHIT-TAX COMPLIANT                 |
|                                                  | P2505NNC3A02        | 1   | ZONE LIGHT, WHIT-TAX COMPLIANT                 |
| NS                                               | P2511BCC-0101-08B-U | 22  | CALL CORD, OXYGEN SAFE, 8 FT                   |
|                                                  | P2516A01-G          | 22  | EQUIPMENT RECEPTACLE (TAA)                     |
| POE-24                                           | P2519NNC2A24        | 1   | POE SWITCH NNC 24-PORT DELTA II                |
|                                                  | P2519NNCFA00        | 2   | MULTI-MODE G8RC FOR TEL-ROOM POE               |
| NS                                               | P2520NNC3A07        | 9   | PB SWITCH, CODE BLUE, WICANCEL                 |
|                                                  | P2520NNC3A08        | 1   | PB SWITCH STAFF EMER WICANCEL                  |
| NS                                               | P2520NNC3B01        | 7   | PULL SWITCH BATH WICANCEL                      |
|                                                  | P2521B02            | 1   | UPS,RACKMOUNTABLE, 2U - NON-SEISMIC            |
| NS                                               | P2594VNC1D01        | 4   | STAFF STATION - STANDARD ROOM STATION W/O CODE |
|                                                  | P2594VNC1D11        | 2   | PATIENT STATION - STANDARD ROOM STATION W/CODE |
| NS                                               | P2599NNC3C00        | 16  | RCB3 ROOM CONTROL BOARD                        |

| System Type = Hillrom CWS Unit Main 1st FL CDD |              |     |                             |
|------------------------------------------------|--------------|-----|-----------------------------|
| 5/16/2024                                      | PNumber      | QTY | Description                 |
| NS                                             | AC19         | 2   | GRS10H WALL MNT W/O HANDSET |
|                                                | P2599NNC3C00 | 2   | RCB3 ROOM CONTROL BOARD     |

| System Type = Hillrom CWS Unit Main 1st FL DENTAL |                |     |                                                |
|---------------------------------------------------|----------------|-----|------------------------------------------------|
| 5/16/2024                                         | PNumber        | QTY | Description                                    |
| NS                                                | AC19           | 13  | INSTALLATION Single-Gang Rough-in              |
|                                                   | P2505NNC3A02-G | 12  | ZONE LIGHT, WHIT-TAX COMPLIANT                 |
| NS                                                | P2505NNC3A02   | 1   | ICON BASED-LIGHT LED ZONE LIGHT                |
| POE-24                                            | P2519NNC2A24   | 1   | POE SWITCH NNC 24-PORT DELTA II                |
|                                                   | P2519NNCFA00   | 2   | MULTI-MODE G8RC FOR TEL-ROOM POE               |
| NS                                                | P2520NNC3A07   | 9   | PB SWITCH, CODE BLUE, WICANCEL                 |
|                                                   | P2520NNC3A08   | 9   | PB SWITCH STAFF EMER WICANCEL                  |
| NS                                                | P2520NNC3B01   | 4   | PULL SWITCH BATH WICANCEL                      |
|                                                   | P2521B02       | 1   | UPS,RACKMOUNTABLE, 2U - NON-SEISMIC            |
| NS                                                | P2594VNC1D01   | 1   | STAFF STATION - STANDARD ROOM STATION W/O CODE |
|                                                   | P2594VNC7C00   | 1   | GRAPHICAL ROOM STATION (GRST) - STAFF          |
| NS                                                | P2599NNC3C00   | 7   | RCB3 ROOM CONTROL BOARD                        |

Main 1st Oncology Drawing



PRELIMINARY  
DRAWING ONLY  
12/07/2023  
DO NOT USE FOR  
CONSTRUCTION  
DRAWN BY: LPK  
CHECKED BY: