


Dental




System Type	Item No.	Unit	Qty	Description
5/16/2024	AC101	1	1	REPLACEMENT Audio Station (GRS, CDS5) Noughs
	AC102	1	1	INSTALLATION Single-gang Noughs
	P5030NCCNAG2-G	64	64	COGE LIGHT W/1-TXA COMPLANT
	P5030NCCNAG2-G	1	1	COGE LIGHT W/1-TXA COMPLANT
	P0511NCCZAG2	2	2	P05 SWITCH NCC-24 PORT DOA B
	P5211NCCZAG2	2	2	MULI-MEDIA FOR HILL-ASBEST
	P5030NCCNAG2	26	26	PS SWITCH COGE BLUE, W/ANCAL
	P5030NCCNAG2	1	1	PS SWITCH COGE EMER W/ANCAL
	P5030NCCNAG2	1	1	PULL SWITCH BATH W/ANCAL
	P5030NCCNAG2	1	1	UPS RACKMOUNTABLE, 2U, NON-SEISMIC
	P5030NCCNAG2	1	1	STAFF STATION - STANDARD ROOM STATION W/O CODE
	P5030NCCNAG2	1	1	GRS10 WALL MNT W/ HANDSET
	P5030NCCNAG2	33	33	ROOM ROOM CONTROL BOARD

Access Type	Access CW2	Unit CW10	
5/26/2024	Phumber		
	A219	2	INSTALLATION Single-Gang Rough-in
	P2505N(NC)A2-G	2	DOME LIGHT, WHT -TAA COMPLIANT
	P2505N(NC)A2-G	1	ZONE LIGHT, WHT -TAA COMPLIANT
	P2519N(NC)A24	2	POE SWITCH NNC 24-PORT DELTA II
	P2519N(NC)A00	2	MULTIMEDIA BRIC FOR HUB-POL POE
	P2520N(NC)B01	1	PULL SWITCH/ATH W/ACCEL
	P2521B02	1	UPS, RACKMOUNTABLE, 2U -NON-SEISMIC
	P2569N(NC)C00	2	ROB3 ROOM CONTROL BOARD

System Type = Hilarom CWS Unit: 2nd FL 2 C Client			
5/16/2024	PNumber	QTY	Description
	AD15	20	INSTALLATION Single-Stage Rough-In
	P2509VNC3J00	1	GRS10L DSK MNT, 4USB
	P2509NNCBA02-C	21	DOHE LIGHT, WHT -TAA COMPLIANT
	P2509NNCBA02-C	2	ZONE LIGHT, WHT -TAA COMPLIANT
	P2509NNCB001	4	PULL SWITCH,BATH W/CANCEL
	P2594VNC1D11	16	PATIENT STATION - STANDARD ROOM STATION W/CODE
	P2509NNC3J00	14	RCB3 ROOM CONTROL BOARD

Ophthalmology



**PRELIMINARY
DRAWING ONLY**
12/07/2023
**DO NOT USE FOR
CONSTRUCTION**

DRAWN BY: LPK
CHECKED BY:

610128		
VETERANS AFFAIR MEDICAL CENTER 521		
700 19TH ST S		
BIRMINGHAM	AL	352331927

BAXTER
NURSE CALL SOLUTIONS
1225 CRESCENT GREEN
SUITE 200
CARY, NC 27518
www.hillrom.com



Revision Date	Revised by	Comments
Revision History		

<u>Drawing Information</u>	
02 DWG QWAQ15097-02 VAMC 521 BIRMINGHAM 2ND FLOOR.dwg	
<u>Sheet Number</u>	<u>Total Sheets</u>
3	12
Drawing Scale	1" = 20'

Project Information
REPLACE NURSE CALL SYSTEM
Proposal #QWAQ15097-01
NURSE CALL LAYOUT

Customer Information