



GENERAL NOTE - HILLROM CWS - USE: Main 3rd Floor			
5/16/2024	Number	QTY	Description
NS	P250VNC300	1	GRS 10H DEK MNT. 4U3B
QD	P250NNC3A02-G	18	DOME LIGHT - WRT-TAA COMPLIANT
PSC-24	P2511NNC2A4	1	POE SWITCH NMC 24-PORT DELTA II
U	P2511RNC7A00	1	MULTI MODE USB FOR BILL-SOM POE
U	P250NNC3A07	10	PR SWITCH COSE BLUE W/CANCELL
PSC-24	P250NNC3B01	3	PULL SWITCH BATH W/CANCELL
PSC-24	P2521B02	1	UPS RACKMOUNTABLE, 2U - NON-SERVMC
QD	P250VNC1D01	1	STAFF STATION - STANDARD ROOM STATION W/O CODE
QD	P250VNC1D11	1	PATIENT STATION - STANDARD ROOM STATION W/O CODE
QD	P250NNC3C00	9	RCB3 ROOM CONTROL BOARD

GENERAL NOTE - HILLROM CWS - USE: Main 3rd Floor			
5/16/2024	Number	QTY	Description
NS	P250VNC300	1	GRS 10H DEK MNT. 4U3B
QD	P250NNC1C00	13	AUDIO STATION BED CONNECTOR
QD	P250NNC3A02-G	31	DOME LIGHT - WRT-TAA COMPLIANT
QD	P250NNC3A02-G	3	DOVE LIGHT - WRT-TAA COMPLIANT
PSC-24	P2511NNC2A4	1	POE SWITCH NMC 24-PORT DELTA II
U	P2511RNC7A00	1	MULTI MODE USB FOR BILL-SOM POE
U	P250NNC3A07	9	PR SWITCH COSE BLUE W/CANCELL
PSC-24	P250NNC3B01	9	PULL SWITCH BATH W/CANCELL
PSC-24	P2521B02	1	UPS RACKMOUNTABLE, 2U - NON-SERVMC
QD	P250VNC1D01	7	STAFF STATION - STANDARD ROOM STATION W/O CODE
QD	P250VNC1D11	13	PATIENT STATION - STANDARD ROOM STATION W/O CODE
QD	P250NNC3C00	24	RCB3 ROOM CONTROL BOARD

GENERAL NOTE - HILLROM CWS - USE: Main 3rd Floor			
5/16/2024	Number	QTY	Description
QD	P250NNC1C00	1	AUDIO STATION BED CONNECTOR
QD	P250NNC3A02-G	1	DOVE LIGHT - WRT-TAA COMPLIANT
QD	P2511NNC2A4	1	PULL SPM DA-MC-DV N2 I.T.F
QD	P250VNC1D11	1	PATIENT STATION - STANDARD ROOM STATION W/O CODE
QD	P250NNC3C00	1	RCB3 ROOM CONTROL BOARD

GENERAL NOTE - HILLROM CWS - USE: Main 3rd Floor			
5/16/2024	Number	QTY	Description
QD	P250NNC3A02-G	1	DOVE LIGHT - WRT-TAA COMPLIANT
QD	P250NNC3B01	1	PULL SWITCH BATH W/CANCELL
QD	P250VNC3B01	2	GRS 10H WALL MNT W/O HANDSET
QD	P250NNC3C00	3	RCB3 ROOM CONTROL BOARD