

BRC 5th Floor
Standard audio on head wall.
Bath
Shower
ASBC mounted beside standard patient station.
Use existing rough ins.

System Type = Hillrom CWS Unit: BRC 5th Floor

5/16/2024	Pnumber	QTY	Description
NS	P2500VNC3C000	1	GRS10H DSK MNT, 4USB
AD	P2505NNC1C00	15	AUDIO STATION BED CONNECTOR
	P2506NNC8A02-G	16	DOME LIGHT, WHT -TAA COMPLIANT
	P2509NNC8A02-G	6	ZONE LIGHT, WHT -TAA COMPLIANT
	P25104C331U-Z105018	15	PILL SPK, DA,MIC,DV,N.2 LT,8'
POC-24	P2519NNC2A24	1	POE SWITCH NNC 24-PORT DELTA II
	P2519NNCFA00	2	MULTI MODE GBIC FOR HILL-ROM POE
	P2520NNC3B01	11	PULL SWITCH,BATH W/CANCEL
	P2520NNC3B02	9	PULL SWITCH, BATH, W/O CANCEL
UPS Rackmount	P2521A02	1	UPS,RACKMOUNTABLE, 2U - SEISMIC APPROVED
SN	P2594VNC1D01	1	STAFF STATION - STANDARD ROOM STATION W/O CODE
SN	P2594VNC1D11	15	PATIENT STATION - STANDARD ROOM STATION W/CODE
RB	P2599NNC3C00	16	RCB3 ROOM CONTROL BOARD

System Type = Hillrom CWS Unit: Main 5th Floor 5B

5/16/2024	Pnumber	QTY	Description
DIS	AD18	1	INSTALLATION:Audio Station (SRS, GRSS) Rough-in
DIS	AD19	4	INSTALLATION:Single-Gang Rough-in
NS	P2500VNC3C000	1	GRS10H DSK MNT, 4USB
AD	P2505NNC1C00	18	AUDIO STATION BED CONNECTOR
	P2506NNC8A02-G	16	DOME LIGHT, WHT -TAA COMPLIANT
	P2509NNC8A02-G	4	ZONE LIGHT, WHT -TAA COMPLIANT
	P25104C331U-Z105018	18	PILL SPK, DA,MIC,DV,N.2 LT,8'
POC-24	P2519NNC2A24	2	POE SWITCH NNC 24-PORT DELTA II
	P2519NNCFA00	2	MULTI MODE GBIC FOR HILL-ROM POE
	P2520NNC3B01	15	PULL SWITCH,BATH W/CANCEL
	P2520NNC3B02	15	PULL SWITCH, BATH, W/O CANCEL
UPS Rackmount	P2521A02	1	UPS,RACKMOUNTABLE, 2U - SEISMIC APPROVED
1:1	P2523C19	2	TRIM PLATE, 3-GANG TO 1-GANG ADAPTOR PLATE
SN	P2594VNC1D01	2	STAFF STATION - STANDARD ROOM STATION W/O CODE
SN	P2594VNC1D11	18	PATIENT STATION - STANDARD ROOM STATION W/CODE
RB	P2599NNC3C00	15	RCB3 ROOM CONTROL BOARD

System Type = Hillrom CWS Unit: Main 5th Floor 5A

5/16/2024	Pnumber	QTY	Description
DIS	AD18	5	INSTALLATION:Audio Station (SRS, GRSS) Rough-in
DIS	AD19	2	INSTALLATION:Single-Gang Rough-in
NS	P2500VNC3C000	1	GRS10H DSK MNT, 4USB
AD	P2505NNC1C00	20	AUDIO STATION BED CONNECTOR
	P2506NNC8A02-G	19	DOME LIGHT, WHT -TAA COMPLIANT
	P2509NNC8A02-G	4	ZONE LIGHT, WHT -TAA COMPLIANT
	P25104C331U-Z105018	20	PILL SPK, DA,MIC,DV,N.2 LT,8'
	P2520NNC3A07	3	PB SWITCH, CODE BLUE, W/CANCEL
	P2520NNC3B01	17	PULL SWITCH,BATH W/CANCEL
	P2520NNC3B02	16	PULL SWITCH, BATH, W/O CANCEL
1:1	P2523C19	6	TRIM PLATE, 3-GANG TO 1-GANG ADAPTOR PLATE
SN	P2594VNC1D01	3	STAFF STATION - STANDARD ROOM STATION W/O CODE
SN	P2594VNC1D11	20	PATIENT STATION - STANDARD ROOM STATION W/CODE
RB	P2599NNC3C00	16	RCB3 ROOM CONTROL BOARD

Main 5th Floor 5A

Main 5th Floor 5B

Main 5th Floor 5A and 5B
Standard audio stations at headwall
Bath
Shower
ASBC at headwall
Use existing rough ins.
Semi private room in 5A requires new rough in per drawing.





NOTICE
The drawings and all information thereon is property of:
BAXTER HEALTHCARE CORPORATION
BATESVILLE, INDIANA
This drawing and information thereon, is confidential and must not be made public, copied, or used to disadvantage of said Baxter Healthcare Corporation
The drawing and all information is subject to return to Baxter on demand.

BAXTER
NURSE CALL SOLUTIONS
1225 CRESCENT GREEN
SUITE 200
CARY, NC 27518
www.hillrom.com

Hillrom™
Hillrom is a part of Baxter

Revision Date	Revised by	Comments

Revision History

Drawing Information
07 DWG QWAQ15097-01 VAMC 521 BIRMINGHAM 5TH FLOOR.dwg
Sheet Number 8
Total Sheets 12
Drawing Scale 1" = 20'
U.L. LISTED

Project Information
REPLACE NURSE CALL SYSTEM
Proposal #QWAQ15097-02
NURSE CALL LAYOUT

Customer Information
610128
VETERANS AFFAIR MEDICAL CENTER 521
700 19TH ST S
BIRMINGHAM AL 352331927