GRS10 location will need to be determined. IT Closet needs to be identified.

System Type = Hillrom CWS Unit: 6th FL

5/16/2024	PNumber	QTY	Description
N5	P2500VNC3C00	1	GRS10H,DSK MNT, 4USB
	P2500BNC8A02-D	6	DOME LIGHT, WHT -TAA COMPLIANT
POE-24	P2519NNC2A24	1	POE SWITCH,NNC 24-PORT DELTA II
	P2520NNC3A07	4	PB SWITCH, CODE BLUE, W/CANCEL
	P2520NNC3A08	2	PB SWITCH,STAFF EMER,W/CANCEL
	P2520NNC3B01	5	PULL SWITCH,BATH W/CANCEL
UPS-RACK MOUNT Battery	P2521A02	1	UPS,RACKMOUNTABLE, 2U - SEISMIC APPROVED
	P2599NNC3C00	5	RCB3 ROOM CONTROL BOARD

N

NOTICE
The drawings and all information thereon is property of:
BAXTER HEALTHCARE CORPORATION
BATESVILLE, INDIANA
This drawing and information thereon, is confidential and must not be made public, copied, or used to disadvantage of said Baxter Healthcare Corporation
The drawing and all information is subject to return to Baxter on demand.

BAXTER
NURSE CALL SOLUTIONS
1225 CRESCENT GREEN
SUITE 200
CARY, NC 27518
www.hillrom.com

Hillrom™
Hillrom is a part of Baxter

Revision Date	Revised by	Comments

Revision History

Drawing Information
08 DWG QWAQ15097-02 VAMC 521 BIRMINGHAM 6TH FLOOR.dwg
Sheet Number: 9
Total Sheets: 12
Drawing Scale: 1" = 20'
U.L. LISTED

Project Information
REPLACE NURSE CALL SYSTEM
Proposal #QWAQ15097-02
NURSE CALL LAYOUT

Customer Information
610128
VETERANS AFFAIR MEDICAL CENTER 521
700 19TH ST S
BIRMINGHAM AL 352331927